

Re Canister Clean Certificate From Chemtex

From: Chemtex-Organics [chemtexpa@sbcglobal.net]
Sent: Monday, April 05, 2010 4:34 PM
To: Agostini, Francis - HLT
Subject: Re: Canister Clean Certificate From Chemtex
THE FINAL PRESSURE WAS -30 INCHES OF Hg WHEN WE CERTIFIED.

----- Original Message -----

From: [Agostini, Francis - HLT](#)
To: [Chemtex-Organics](#)
Cc: [Hoyt, Dan - HLT](#)
Sent: Monday, April 05, 2010 3:35 PM
Subject: RE: Canister Clean Certificate From Chemtex

Thanks for sending the clean certificate. I require a detailed certificate that includes the pressure at which each canister was certified. Please send the more detailed certificate as soon as possible.

Thanks,

Francis Agostini, Graduate Engineer
Bureau of Air Quality Control
Environmental Health Division
Health and Human Services Department
City of Houston
7411 Park Place Blvd.
Houston, TX 77087
832.393.5617 (office)
713.483.4914 (cell)
713.640.4341 (fax)

I count. You count. Houston counts.
Answer the Census
www.houstoncounts.org

From: Chemtex-Organics [mailto:chemtexpa@sbcglobal.net]
Sent: Monday, April 05, 2010 3:14 PM
To: Agostini, Francis - HLT
Subject: Re: Canister Clean Certificate From Chemtex

----- Original Message -----

From: [Agostini, Francis - HLT](#)
To: 'CHEMTEX'
Cc: [Hoyt, Dan - HLT](#)
Sent: Friday, March 26, 2010 11:13 AM
Subject: City of Houston DIAL canisters

Thanks for bringing back to our offices the canisters 4569, 4570, and 1350. Please send me the cleaning and certification documentation by mail or email as soon as possible.

In addition, thank you for picking up canister 1353 for analysis. For your reference, PID readings during sample were around 800 ppbv.

Thanks,

Francis Agostini, Graduate Engineer
Bureau of Air Quality Control
Environmental Health Division

Re Canister Clean Certificate From Chemtex

Health and Human Services Department
City of Houston
7411 Park Place Blvd.
Houston, TX 77087
832.393.5617 (office)
713.483.4914 (cell)
713.640.4341 (fax)

I count. You count. Houston counts.

Answer the Census

www.houstoncounts.org

From: CHEMTEX [mailto:chemtexpa@sbcglobal.net]

Sent: Friday, March 05, 2010 2:11 PM

To: Agostini, Francis - HLT

Subject:

We compleated the analysis .do you what us to clean your TO-CANs?.please let us know.

Thanks,

Sudheer.

CHEMTEX

3082 25th St.

PORT ARTHUR, TX 77642.

Phone: (409) 983-4575

Fax: (409) 982-1522

e-mail: chemtexpa@sbcglobal.net